



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/669,240	09/23/2003	Hiroshi Koguchi	2102487-991280	3543
26379 7590 05/09/2007 DLA PIPER RUDNICK GRAY CARY US, LLP 2000 UNIVERSITY AVENUE E. PALO ALTO, CA 94303-2248			EXAMINER WANG, U LUN	
			ART UNIT 2616	PAPER NUMBER
			MAIL DATE 05/09/2007	DELIVERY MODE PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

sf

Office Action Summary	Application No.	Applicant(s)	
	10/669,240	KOGUCHI, HIROSHI	
	Examiner	Art Unit	
	U-Lun Wang	2616	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 23 September 2003.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-14 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1 and 9 is/are rejected.
- 7) ☒ Claim(s) 2-8 and 10-14 is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
 Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
 Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☒ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☒ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| 3) <input checked="" type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08)
Paper No(s)/Mail Date <u>6/3/2005, 9/23/2003</u> . | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Claim Rejections - 35 USC § 102

1. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(e) the invention was described in (1) an application for patent, published under section 122(b), by another filed in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

2. Claims 1 and 9 are rejected under 35 U.S.C. 102(e) as being anticipated by Nomura (US Patent 7,035,272 B2).

For claim 1, Nomura discloses a device capable of correcting time stamp value comprising: an original time counter (Col. 1, lines 42 – 43: Each device connected to the IEEE-1394 serial bus has a cycle time register. Note that the cycle time register is a time counter) configured to generate and output an original time value; a subtracter configured to receive a network time value from the outside (Fig. 14, S83: the cycle count value of the extracted time stamp. Note that the time stamp is a network time value from outside) and the original time value from the original time counter (Fig. 14, S83: the cycle count value of the cycle time register), subtract the original time value from the network time value (Fig. 14, S83: Calculate the difference. Note that this implies there is a subtracter which calculates the difference), and output the difference as a cycle-time difference value (Fig. 14, S95: DiffCycleCount); and a first adder configured to receive an original time-stamp value generated from the original time

Art Unit: 2616

value and the cycle-time difference value from the subtracter, add the cycle-time difference value to the original time-stamp value (Fig. 14, S95: Add DiffCycleCount to the original time stamp), and output the sum as a time-stamp value (Fig. 14, S95: And store the result in "Sync Time" of the extracted CIP).

For claim 9, Nomura discloses all the subject matter (see rejection of claim 1 above).

Allowable Subject Matter

3. Claims 2 – 8 and 10 – 14 are objected to as being dependent upon a rejected base claim, but would be allowable if rewritten in independent form including all of the limitations of the base claim and any intervening claims.

Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to U-Lun Wang whose telephone number is (571) 270-1140. The examiner can normally be reached on Monday - Friday, 7:30 AM - 5 PM.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Chau Nguyen can be reached on 571-272-3126. The fax

Art Unit: 2616

phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

UW/
4/23/2007



CHAU NGUYEN
SUPERVISORY PATENT EXAMINER
TECHNOLOGY CENTER 2600